



TEST REPORT

98, Gyooyukwon-ro, Gwacheon-si, Gyeonggi-do, 13810, Korea

TEL 82-2-2164-0011

FAX 82-2-2634-1008

Report No : TBK-2020-001054

Receipt Date : 2020.02.21.

Representative : Eric Choi

Test Completion Date : 2020.04.03.

Company name : Haenim CO.,LTD

Address : 15, Saryeom-ro 21beon-gil, Seo-gu, Incheon, Republic of Korea

Sample name : Haenim UV LED Bottle Sterilizer

Test Results

TEST ITEM	UNIT	SAMPLE	RESULT	TEST METHOD
Virucidal test[Human Influenza A (H1N1)]	Log reduction	-	3.12	Client Provided Test Method

* Test conditions

- UV LED mode (operation time: 10 minutes)

** Test virus

- Human Influenza A (H1N1) (Strain: KUMC-76), KBPV-VR-76

*** Host cells

- MDCK (NBL-2), ATCC CCL-34

**** Log reduction (LR)

- LR = LU - LT

- LU = The titer recovered from the virus control specimen (untreated)

- LT = The titer recovered from virucidal test specimen (treated)

***** Interpretation of the results

- log reduction > 1: > 90 %

- log reduction > 2: > 99 %

- log reduction > 3: > 99.9 %

- log reduction > 4: > 99.99 %

- log reduction > 5: > 99.999 %

- Usage of Report : QUALITY CONTROL

- Next Page -

Hyun Ho Kim

Prepared by Hyun Ho Kim

Tel : 02-2092-3956

Jin-Sik Cho

Reviewed by Jin-Sik Cho

Tel : 1577-0091(ARS ①-④)

2020.04.03

Korea Testing & Research Institute

President *Kwon Oh-jung*



QR Code for forgery

Registered copy date : Apr.08.2020

Page : 1 of 2